



DECLARATION OF CONFORMITY

Regarding Medical Device Directive(93/42/EEC)
including Directive 2007/47/EC

Manufacturer: Guangdong Winsun Personal Care Products Co.,Ltd
Address: NO 1 Guangxing Road,Xiqiao Sci-Tec Industrial Park,Xiqiao
Town,Nanhai District,Foshan City,Guangdong Province P.R.China

EC Representative: SUNGO Europe B.V.
Address: Olympisch Stadion 24, 1076DE Amsterdam Netherlands

Product Name: Disposable Medical Face Mask
Specification: 17.5cm*9.5cm, 14.5cm*9.5cm

Classification: Class I (MDD, Annex IX)

Conformity Assessment

Procedure: Annex VII of Medical Device Directive(93/42/EEC)

We herewith declare that the above-mentioned products meet the requirements of
Medical Device Directive (93/42/EEC) including Directive 2007/47/EC and the following
harmonized standards.

EN ISO 14971:2012

EN ISO 15223-1:2010

EN 1041:2008+A1:2013

ISO 10993-1:2010

EN ISO 10993-5:2009

EN ISO 10993-10:2013

EN 14683:2010

Signature: 
Name: Foshan Xiqiao Yaru GM

Date: 2020.5.8
Place: Guangdong, China

*On behalf of SUNGO Europe office, I confirmed we are
EU-REP of the company who issue this document.*



Authorized Signature (S)

Certificate CN20/42094

The management system of

Guangdong Winsun Personal Care Products Co., Ltd.

NO. 1 Guangxing Road, Xiqiao Sci-Tec Industrial Zone, Xiqiao Town,
Nanhai District, Foshan City, Guangdong Province, 528211, P.R. China

has been assessed and certified as meeting the requirements of

ISO 13485:2016 EN ISO 13485:2016

For the following activities

Design and manufacture of single use non-sterile medical face mask

This certificate is valid from 17 June 2020 until 16 June 2023
and remains valid subject to satisfactory surveillance audits.
Re certification audit due before 2 June 2023
Issue 1. Certified since 17 June 2020

Authorised by



SGS United Kingdom Ltd
Rossmore Business Park Ellesmere Port Cheshire CH65 3EN UK
t +44 (0)151 350-6666 f +44 (0)151 350-6600 www.sgs.com

HC SGS 13485 2016 0118

Page 1 of 1





认证证书

证书号: USA19Q42833R1M

兹证明

广东昱升个人护理用品股份有限公司

统一社会信用代码: 914406007820115710

注册地址: 佛山市南海区西樵镇西樵科技工业园广兴路1号

生产地址: 广东省佛山市南海区西樵镇西樵科技工业园广兴路1号车间一

办公地址: 广东省佛山市南海区西樵镇西樵科技工业园广兴路1号

质量管理体系符合标准

ISO 9001:2015

质量管理体系适用范围

纸尿裤、纸尿片的生产(该公司许可证范围内)

初次发证日期 2010年04月22日


证书颁发日期 2019年07月17日

证书有效期至 2022年07月16日

签发: **吴凤茹**



北京东方纵横认证中心有限公司



获证组织必须定期接受监督审核并经审核合格此证书方继续有效。本证书信息可在北京东方纵横认证中心有限公司网站(www.eacc.com.cn)和国家认证认可监督管理委员会官方网站(www.cnca.gov.cn)上查询,也可扫描右下角的二维码查询。

认证机构地址: 北京市通州区中关村科技园区通州园金桥科技产业基地景盛南四街17号121号楼一层101102





Certificate

Certificate No.: USA19Q42833R1M

This is to certify that the Quality Management System of
**GUANGDONG YUSHENG PERSONAL
NURSING SUPPLIES CO., LTD.**

Unified Social Credit Code: 914406007820115710

Registered/Office Address: No. 1 Guangxing Road, Xiqiao Technology Industry Park,
Xiqiao Town, Nanhai District, Foshan, Guangdong

Production Address: Workshop 1, No. 1 Guangxing Road, Xiqiao Technology Industry Park,
Xiqiao Town, Nanhai District, Foshan, Guangdong

Has been audited to conform to the following Quality Management System standard
ISO 9001:2015

This Quality Management System is valid for the
Production of diapers and diapers
(within the scope of the company's license)

Date of initial issuance: Apr. 22, 2010

Date of issuance: Jul. 17, 2019

Date of expiry: Jul. 16, 2022

Issued by: *Wu Fengyu*



Beijing East Allreach Certification Center Co., Ltd.



The certificate will remain valid only if the certified organization accepts surveillance audits at regular intervals and is audited to be qualified. The information of this certificate is available at EACC website (www.eacc.com.cn/) and CNCA's official website (www.cnca.gov.cn/), and it is also available by scanning the QR Code in the lower right corner.

EACC address 1st Floor, No. 121 building, No. 17 Jingshengjianshi Street, Jinqiao Science & Technology Industrial Base, Tongzhou Park of Zhongguancun Science & Technology Zone, Tongzhou District, Beijing 101102



ISO 9001

ISO 9001

Test Report

SL52025258960201TX

Date: June 19, 2020

Page 1 of 4

GUANGDONG WINSUN PERSONAL CARE PRODUCTS CO., LTD
NO 1 GUANGXING ROAD, XIQIAO SCI-TEC INDUSTRIAL PARK, XIQIAO TOWN, NANHAI DISTRICT,
FOSHAN CITY, GUANGDONG PROVINCE

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Disposable Medical Face Mask(Claimed Type IIR)
Composition : (A) Spunbond nonwoven, meltblown nonwoven, elastic earloop, nose strip etc
Sample Color : (A) Blue/White
Lot No./Batch No. : Not provided

Proposed Care Instruction : -

Sample Receiving Date : May 28, 2020
Testing Period : May 28, 2020 - Jun 19, 2020
Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Test Performed : Selected test(s) as requested by applicant

Comment:

| | |
|---|----------|
| EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods | (A) |
| Clause 5.2 Performance Requirement | |
| Clause 5.2.2 Bacterial filtration efficiency (BFE) | M |
| Clause 5.2.3 Breathability | M |
| Clause 5.2.4 Splash Resistance | M |
| Clause 5.2.5 Microbial Cleanliness | M |
| Clause 5.2.6 Biocompatibility | EXCLUDED |

Remark: M=Meet EN 14683:2019+AC:2019 Performance Requirement (Type IIR)
F=Below EN 14683:2019+AC:2019 Performance Requirement (Type IIR)

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu / Hailian Xuan (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced or copied in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing inspection report & certificate, please contact us at china@sgs.com or visit CN.Questia@sgs.com

Building No. 888, Yangtze Road, Xuhui District, Shanghai, China 200023 | (86-21) 61402886 | (86-21) 64856703 | www.sgs.com.cn
中国·上海·徐汇区宜山路888号3号楼 邮编 200023 | (86-21) 61402886 | (86-21) 64856703 | sgs.china@sgs.com

Test Report

SL52025258960201TX

Date: June 19, 2020

Page 2 of 4

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.2 Bacterial filtration efficiency (BFE)

(EN 14683:2019+AC:2019 Annex B)

Sample: A

Conditioning Parameters : Minimum of 4 hours at 21±5°C and 85±5% R.H.
 Dimensions of test specimen : ~172 mm x 152 mm
 Test Area : ~60 cm²
 Test Side : Inside
 Flow Rate : 28.3 l/min
 Positive Control Average : 2914.5 CFU
 Negative Monitor Count : < 1 CFU

| | 1# | 2# | 3# | 4# | 5# |
|----------|------|------|------|------|------|
| (BFE), % | 99.9 | 99.9 | 99.9 | 99.9 | 99.9 |

Remark: Performance Requirement: Type I ≥95%, Type II ≥98%, Type IIR ≥98%

Clause 5.2.3 Breathability

(EN 14683:2019+AC:2019 Annex C)

Sample: A

Test number and location : 5 random areas for each specimen (face mask)
 Conditioning Parameters : Minimum of 4 hours at 21±5°C and 85±5% R.H.
 Test Area : 4.9 cm²
 Flow Rate : 8 l/min

| | 1# | 2# | 3# | 4# | 5# |
|---|----|----|----|----|----|
| Differential pressure ΔP (Pa/cm ²) | 31 | 31 | 32 | 33 | 32 |

Remark: Performance Requirement: Type I <40 Pa/cm², Type II <40 Pa/cm², Type IIR <60 Pa/cm²



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
 Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-159) 2027 1841 or email: CN.Document@sgs.com

Clause 5.2.4 Splash Resistance

(ISO 22609 :2004, Pressure 16.0 kPa)

Sample: A

| Penetration on inside surface | | | | | | | |
|-------------------------------|------|------|------------|------|------|------|------|
| 1# | 2# | 3# | 4# | 5# | 6# | 7# | 8# |
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | Pass |
| 9# | 10# | 11# | 12# | 13# | 14# | 15# | 16# |
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | Pass |
| 17# | 18# | 19# | 20# | 21# | 22# | 23# | 24# |
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | Pass |
| 25# | 26# | 27# | 28# | 29# | 30# | 31# | 32# |
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | Pass |
| Number of Pass: | | | 32 | | | | |
| Overall result: | | | Acceptable | | | | |

Remark:

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: ≥ 16.0 kPa
- 2) Distance of the medical face mask target area surface to the tip of cannula is 300 ± 10 mm.
- 3) Condition and Test temperature $(21 \pm 5)^\circ$ C, relative humidity $(85 \pm 10)\%$
- 4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results

Clause 5.2.5 Microbial Cleanliness

(EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

Sample: A

| | 1# | 2# | 3# | 4# | 5# |
|-------|----|----|----|----|----|
| CFU/g | 1 | 1 | 1 | <1 | 1 |

Remark: Performance Requirement: Type I ≤ 30 CFU/g, Type II ≤ 30 CFU/g, Type IIR ≤ 30 CFU/g



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.cn/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.cn/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only. Attention: To check the authenticity of testing inspection report & certificate, please contact us at sgs@sgs.com or CN_Support@sgs.com.